



Certificate of Performance

Product Name: ZearaStar™
Product No.: COIAC4000
Lot Number: ZE659 /101017
Expiry Date: 01.11.2019
Storage Conditions: 2-8° C, do not freeze
Code of Experiment: AS750-809 /091017
Testing Method: ZearaStar™ columns were tested at 75 ng and 500 ng of zearalenone.
Loading Conditions: 20 ml 15% methanol/0.01M PBS, pH 7.2 at 1ml/min
Washing with 20 ml deionized water
Elution with 1.5 ml methanol

Results:

Toxin	75 ng	500 ng
Zearalenone	95% +/- 5%	94% +/- 2%

According to these results, ZearaStar™ Lot Number ZE659 passed the internal quality control procedures and is released for sale.

Approved By: Lilian Kuster

Date: 11-October-2017